


EinScan-SP 3D Scanner

EinScan-SP 3D Scanner	
	
Tool Type: "3D Scanner"	
Location: "Innovations Workshop"	
Supervisor	Tool Lead
David Bothman	"WW Name"
(805) 893-4125	(###) ###-####
bothman@cnsi.ucsb.edu	"WW Email"
Description: "3D Scanner"	
Manufacturer: "EinScan"	

About

The SHINNG3D EinScan-SP 3D Scanner is located within Fume Hood #1 under the blackout curtains. A part can be placed on the central turn table which will rotate as the scanner projects light onto the object being scanned. Transparent, reflective, or very dark surfaces are inherently difficult to scan, coating the surface with powder may aid the quality of the scan. Calibration of the 3D scanner can aid in scan accuracy, and should be done before scanning the first object.

Training Documentation

[EinScan SOP](#)

Detailed Specifications

Part Volume Requirements: must be larger than 30 x 30 x 30 mm and no larger than 250 x 250 x 250 mm

Part Weight Requirements: must be less than 5 kg

Scan Resolution: .17-.2 mm

Safety Concerns

Do not look directly into the scanners light projector. it bright

Operating Procedures

Insert Text Here!

Reference Documentation

[einscan-sp_guidebook.pdf](#)

[sp-user_manual.pdf](#)

<https://www.youtube.com/watch?v=nnU-WNGqDRI&list=PLtJFjqd-EnwvqeVq8h7Jw1bTtG8a0VwQB&index=3&t=0s>

From:

<https://bpm-wiki.cnsi.ucsb.edu/> - NSF BioPACIFIC MIP Wiki

Permanent link:

<https://bpm-wiki.cnsi.ucsb.edu/doku.php?id=einscan-sp&rev=1598901407>

Last update: **2020/08/31 19:16**

