
# **Zeiss Inverted Microscope**



#### About

Inverted microscope Axio Observer 7 ACR for transmitted-light brightfield LED, phase contrast, differential interference contrast, fluorescence with Colibri 7; equipped with ACR function, Scanning Stage STEP and Docking station

## **Detailed Specifications**

Last update: 2021/07/23 19:41	inverted microscope https://hom.wilki.enci.usch.edu/delku.php?id_inverted_microscopeS.rov_1627060212
Last upuate: 2021/07/25 19:41	inverted_microscope https://bpm-wiki.cnsi.ucsb.edu/doku.php?id=inverted_microscope&rev=1627069313

Objective	Magnification	lmm	NA	Ph	FWD	Resolution (µm/px)	Field of View (µm)
10X EC Plan-Neofluar Ph1- Air	10	AIR	0.3	1	5.2	0.586	1,125 x 713
20X Plan-Apochromat M27	20X	AIR	0.8	1	0.55	0.293	563 x 356
20X LD-Neofluar Korr Ph2 M27 Air	20X	AIR	0.4	2	7.4 – 8.4	0.293	562 x 356
40X LD Plan-Neofluar Korr Ph2 M27-Air	40X	AIR	0.6	2	3.3 – 2.5	0.1465	281 x 178
40X LD LCI Plan-Apochromat Imm Korr DIC M27 Water	40X	WATER	1.2	2	0.41	0.1465	281 x 178
63X Plan-Apochromat Oil DIC M27 Oil	63X	OIL	1.4	2	0.19	0.093	178 x 113
FWD: Free Working Distance							

FWD: Free Working Distance NA: Numerical Aperature Imm: Immersion media

Ph: Phase contrast

Position	Filter Set	Exitation (nm)	Emmision (nm)	Beamsplitter
1	38 HE GFP	BP 470/40	BP 525/50	FT 495
2	43 HE DsRED	BP 550/25	BP 605/70	FT 570
3	50 CY5	BP 640/30	BP 690/50	FT 660
4	90 HE D/G/C3	BP 385/30 BP 469/38 BP 555/30	QBP 425/30 QBP 514/30 QBP 592/25	QBS 405 QBS 493 QBS 575
5	96 HE BEP	BP 631/33 BP 390	QBP 709/100 BP 450	QBS 653 FT 420
6	POL TL	570	51 430	11 420

**Safety Concerns** 

Read the manufactures manual before first use. If the Zeiss inverted microscope acts in a way that is not described by the manual, turn off the printer and contact Zeiss.

- Never place your finger near the machine until all parts have stopped moving. Moving parts can cause serious injury
- Never clean or service the microscope while it is on
- The microscope uses UV light to excite some dyes. Never look directly at LED light nor expose skin. Serious injury may result from exposure
- Disassembling the microscope may cause an electric shock or damage to the instrument. Do not disassemble any parts of the microscope not mentioned in the instruction manual. In case of a problem with the microscope

#### **Operating Procedures**

Insert Text Here!

NSF BioPACIFIC MIP Wiki - https://bpm-wiki.cnsi.ucsb.edu/

#### **Reference Documentation**

 $zen\_2\_blue\_edition\_-software\_guide.pdf$ 

### **Training Documentation**

From: https://bpm-wiki.cnsi.ucsb.edu/ - NSF BioPACIFIC MIP Wiki

Permanent link: https://bpm-wiki.cnsi.ucsb.edu/doku.php?id=inverted\_microscope&rev=162706931

Last update: 2021/07/23 19:41

