

# LumiDox Gen II LED Plate

|                                                          |                     |                                       |                                |                            |
|----------------------------------------------------------|---------------------|---------------------------------------|--------------------------------|----------------------------|
| <b>LumiDox Gen II LED Plate</b>                          | <a href="#">png</a> | Tool Type: LED Plate for Cell Culture | Manufacturer: Analytical Sales | Location: Elings Hall 2436 |
| Principal Scientist                                      |                     |                                       |                                |                            |
| Juan Manuel Urueña                                       |                     |                                       |                                |                            |
| <a href="mailto:jmurueña@ucsb.edu">jmurueña@ucsb.edu</a> |                     |                                       |                                |                            |

## About

The LumiDox Gen II LED Plate is an innovative tool designed for precise light delivery to 96-well cell culture plates. This system utilizes high-intensity LEDs arranged to provide uniform, controlled light exposure across all wells, making it ideal for light-activated cellular experiments. The plate allows for customizable light intensities and wavelengths, supporting a variety of experimental conditions. The LumiDox Gen II's user-friendly interface and robust design enable consistent and repeatable illumination, advancing research in areas such as optogenetics, photobiology, and cell signaling.

## Detailed Specifications

- LED technology: High-intensity LEDs
- Compatible formats: 96-well plates
- Wavelength range: 380-780 nm
- Power control: Adjustable power settings
- Uniformity: High across all wells
- Temperature control: Integrated cooling system

## Safety Concerns

The LumiDox Gen II LED Plate produces high-intensity light, which requires the following safety precautions:\

- Always inspect the LEDs before use to ensure no damage or malfunction.\
- Do not stare directly at the light sources during operation.\
- Use protective eyewear rated for specific wavelengths, especially when working with high-intensity light in the UV range.\
- Ensure that all users are familiar with proper operation procedures and safety protocols outlined by the manufacturer.

From: <https://bpm-wiki.cnsi.ucsb.edu/> - NSF BioPACIFIC MIP Wiki

Permanent link: [https://bpm-wiki.cnsi.ucsb.edu/doku.php?id=led\\_plate&rev=1728526506](https://bpm-wiki.cnsi.ucsb.edu/doku.php?id=led_plate&rev=1728526506)

Last update: **2024/10/10 02:15**

