

# LUMEN X

<b>LUMEN X</b>	
<input checked="" type="checkbox"/>	
<b>Tool Type:</b> "3D Printer"	
<b>Location:</b> "Elings Hall 2436"	
<b>Supervisor</b>	<b>Tool Lead</b>
Juan Manuel Urueña	"WW Name"
jmuruena@ucsb.edu	"WW Email"
<b>Description:</b> "Solution Mask Liquid Lithography (SMaLL)"	
<b>Manufacturer:</b> "CELLINK Life Sciences"	

## About

SMaLL is capable of creating unique structures with mechanically and chemically distinct regions. By using tunable, visible wavelength photoswitches, in combination with orthogonal crosslinking reactions, a single resin formulation can be employed for the facile, one-step printing of complex, bioinspired structures. The fabrication of soft joints and brick-and-mortar architectures showcase the utility of SMaLL for printing all-polymer composites with well-defined regions for different mechanical and chemical properties.

## Detailed Specifications

- wavelength: 385 nm
- luminous flux: 2,500 lumens

Power in mW/cm<sup>2</sup> at certain intensity percentages

25%	30%	35%	40%	45%	50%	55%	60%	65%	70%	75%
8.12	10.36	13.30	15.83	18.24	20.58	22.58	24.37	26.42	28.31	29.85

## Safety Concerns

Read the manufacturer's manual before first use. If the BIO X Cellink acts in a way that is not described by the manual, turn off the printer and contact CELLINK.

- Never place your finger near the machine until all parts have stopped moving. Moving parts can cause serious injury
- Never clean or service the printer while it is on
- The printer uses UV light for sterilization and curing. Never look directly at UV light nor expose skin. Serious injury may result from exposure

- The printer has heated surfaces that can reach temperature up to 250 degrees Celsius. Never touch these surfaces when using the heating function. Allow things to cool before opening, touching the printer
- Always ensure that equipment is correctly mounted before use. Improperly mounted print beds, printheads, cartridges, cables liquid spouts and air spouts can be dangerous. If any equipment appears damaged, turn off the printer, unplug all connections and contact CELLINK.

---

## Operating Procedures

Insert Text Here!

---

## Reference Documentation

[creationworkshop.pdf](#)

[small\\_projector\\_manual.pdf](#)

---

## Training Documentation

[biox\\_manual.pdf](#)

---

From:  
<https://bpm-wiki.cnsi.ucsb.edu/> - **NSF BioPACIFIC MIP Wiki**



Permanent link:  
[https://bpm-wiki.cnsi.ucsb.edu/doku.php?id=lumen\\_x&rev=1623805232](https://bpm-wiki.cnsi.ucsb.edu/doku.php?id=lumen_x&rev=1623805232)

Last update: **2021/06/16 01:00**