

Trotec Speedy 100

Trotec Speedy 100



Tool Type: "Laser Cutter"

Location: "Microfluidics lab"

Supervisor	Tool Lead
David Bothman	Jeran Bruce & Vedad Bassari
(805) 893-4125	(424) 610-6312 & (818) 942-5523
bothman@cnsi.ucsb.edu	jrbruce@ucsb.edu & vedad@ucsb.edu

Description: "Bright Light Cutter"

Manufacturer: "Trotec"

About

One of two laser cutters, the Trotec is located in the Innovations Workshop above its fume extractor. Both laser cutters utilize CorelDraw as a 2D sketch manager which is then imported into Trotec's specific cutting software. CorelDraw can be used to create the 2D sketch, however importing a DXF file or PDF into CorelDraw from Solidworks or other CAD packages is preferred due the CAD packages integrated features and functions.

Training Documentation

[Laser Cutter Training SOP](#)

Detailed Specifications

Work Area: 910 x 305 mm

Max Workpiece Height: 125 mm

Laser Power: 10-120 Watts

Safety Concerns

- Looking directly into the laser can cause retinal damage.
- Confirm that the fume collection system is running whenever the laser is cutting or engraving.
- See list of approved materials for laser cutting, some require nitrogen gas if flammable, or could release chlorine gas if cut.
- NO NOT CUT NON APPROVED MATERIALS INCLUDING METALS.
- Laser lenses must be cleaned within ONE WEEK of time of use. If lenses has not been cleaned, clean before use to avoid damaging lenses.

Reference Documentation

[Marking Tape/Paint](#)

[Operation Manual](#)

[Service Manual](#)

[Plastic Processing Guide](#)

[Job Control Software Manual](#)

[Laser cutting data](#)

From:
<https://bpm-wiki.cnsi.ucsb.edu/> - **NSF BioPACIFIC MIP Wiki**

Permanent link:
https://bpm-wiki.cnsi.ucsb.edu/doku.php?id=trotec_speedy&rev=1598899766

Last update: **2020/08/31 18:49**

