Keyence Microscope Training SOP

Last edited: Furst (11/03/20)	
Instructor:	
Date:	Attendees:
Name Group or Company Signature 1 2 3 4 5 6	

Overview:

- This training provides an introduction to using and operating the Keyence microscope including:
 - Changing Objectives
 - 0-50x and 20-200x
 - 100-1000x
 - Head tilt
 - Stage movement
 - Software
 - Lighting
 - Measure
 - Depth up
 - Image quality
 - Optimize
 - Stitching
 - Safety
 - General microscope use
 - Maintenance

Safety

Microscope Setup

At the microscope:

- 1. Changing objectives
 - 1. Press the button on the top of the objective (right above the Y axis of rotation screw)
 - 2. Twist the camera body counterclockwise and carefully lift directly off. Set carefully on its side to avoid damaging the lenses in the camera assembly

- 3. Loosen the large plastic wingnut screw located above the axis of rotation adjustment on the objective. (DO NOT ADJUST AXIS OF ROTATION SCREWS)
- 4. Lift the objective straight up off the mounting post on the microscope body and place carefully in drawer located under Keyence computer.
- 5. Remove the desired objective and lower over the mounting post on microscope body.
- 6. Gently snug up the wing nut screw to clamp on the flat of the mounting post
- 7. Gently lower the camera assembly onto the top of the objective
- 8. Once lined up, rotate camera assembly clockwise until a sharp click is heard
- 2. Tilting the head
- 3. Changing lighting elements
- 4. moving/positioning the stage

At the Computer:

- 1. Initialize the XY stage
- 2. Software lighting adjustment throw in image quality as well
- 3. Depth up
- 4. Measure
- 5. Stitching
- 6. saving images

Maintenance

Keyence microscope Quick Review

Tool Lead:

Contact: andrewfurst@ucsb.edu

Safety Concern

Safe Operation Procedures Review

Post Processing

Maintenance

From:

https://bpm-wiki.cnsi.ucsb.edu/dokuwiki/ - NSF BioPACIFIC MIP Wiki

Permanent link:

https://bpm-wiki.cnsi.ucsb.edu/dokuwiki/doku.php?id=keyence_sop&rev=1604596622

Last update: 2020/11/05 17:17

