


# Shimadzu Nexera Analytical HPLC-MS

**Shimadzu Nexera HPLC-MS**



**Tool Type:** Analytical Characterization

**Location:** Elings Hall 2411

Principal Scientist	Training and Operations Lead
Morgan Bates	Zachary Nett
morganbates@ucsb.edu	zjnett@ucsb.edu

## About

The Shimadzu Nexera Analytical HPLC system enables the identification and quantification of complex mixtures, while also supporting method development for sample purification on both preparative and semi-preparative scales. Equipped with a 6-way switch valve, the system streamlines the selection between four distinct columns, each offering unique surface chemistry and retention properties. These include C18 columns with 100 and 300 Å pore sizes, a C4 column, and a HILIC column. The system's design also allows for easy eluent interchanges, ensuring flexible adaptation to various purification and characterization requirements. Additionally, it features a dedicated PDA and a single quad LCMS-2050 mass spectrometer, which is capable of both ESI and APCI ionization and covers a mass range of 10-2000 m/z.

## Solvent Guidelines and Sample Holder Options

### Solvent Compatibility

SP Genevac evaporators are robust systems designed to be resistant to the most common solvents and acids used within the chemistry laboratory, including TFA and HCl. **However, the Genevac is not compatible with the following:**

- Carbon disulfide
- Diethyl ether
- 1,2-Dimethoxyethane
- 2-Ethoxyethanol
- Hydrogen cyanide
- Pentane
- Triethylsilane

## Sample formats

Sample holders for the following formats are available:

- **Vials:** 40, 30, 20, and 8-mL sizes
- **Test tubes:** 16, 18, and 25-mm diameter and up to 150 mm length
- **Centrifuge tubes:** 50-mL size

---

From:  
<https://bpm-wiki.cnsi.ucsb.edu/dokuwiki/> - NSF BioPACIFIC MIP Wiki

Permanent link:  
[https://bpm-wiki.cnsi.ucsb.edu/dokuwiki/doku.php?id=shimadzu\\_nexera\\_analytical\\_hplc&rev=1729108211](https://bpm-wiki.cnsi.ucsb.edu/dokuwiki/doku.php?id=shimadzu_nexera_analytical_hplc&rev=1729108211)

Last update: **2024/10/16 19:50**

