

Trotec Speedy 100

Trotec Speedy 100



Tool Type: "Laser Cutter"

Location: "Microfluidics lab"

Supervisor	Tool Lead
David Bothman	Jeran Bruce & Vedad Bassari
(805) 893-4125	(424) 610-6312 & (818) 942-5523
bothman@cnsi.ucsb.edu	jrbruce@ucsb.edu & vedad@ucsb.edu

Description: "Bright Light Cutter"

Manufacturer: "Trotec"

About

One of two laser cutters, the Trotec is located in the Innovations Workshop above its fume extractor. Both laser cutters utilize CorelDraw as a 2D sketch manager which is then imported into Trotec's specific cutting software. CorelDraw can be used to create the 2D sketch, however importing a DXF file or PDF into CorelDraw from Solidworks or other CAD packages is preferred due the CAD packages integrated features and functions.

Training Documentation

[Laser Cutter Training SOP](#)

Detailed Specifications

Work Area: 910 x 305 mm

Max Workpiece Height: 125 mm

Laser Power: 10-120 Watts

Safety Concerns

Looking directly into the laser can cause retinal damage. Confirm that the fume collection system is running whenever the laser is cutting or engraving. See list of approved materials for laser cutting, some require nitrogen gas if flammable, or could release chlorine gas if cut. **NO NOT CUT NON APPROVED MATERIALS INCLUDING METALS.** Laser lenses must be cleaned within ONE WEEK of time of use. If lenses has not been cleaned, clean before use to avoid damaging lenses.

Reference Documentation

[Marking Tape/Paint](#)

[Operation Manual](#)

[Service Manual](#)

[Plastic Processing Guide](#)

[Job Control Software Manual](#)

[Laser cutting data](#)

From:

<https://bpm-wiki.cnsi.ucsb.edu/dokuwiki/> - **NSF BioPACIFIC MIP Wiki**

Permanent link:

https://bpm-wiki.cnsi.ucsb.edu/dokuwiki/doku.php?id=trotec_speedy&rev=1598899733

Last update: **2020/08/31 18:48**

